

Compatibility of Current Sensors for Qualistar and PEL

| | CA8220 | CA8331 CA8333 CA8336 | CA8436 | CA8345 | PEL51 PEL52 | PEL102 PEL103 PEL104 | PEL106 |
|--------|--------|----------------------------|-----------|--------|-------------|----------------------------|-----------|
| MN93 | x | x | x (1) | x | x | x | x (1) |
| MN93A | x | x | x (1) | x | x | x | x (1) |
| Mini94 | | | | x | x | | |
| C193 | x | x | x (1) | x | x | x | x (1) |
| MA194 | x | x | x (1) | x | x | x | x (1) |
| MA196 | | | x | | | | x |
| A193 | x | x | x (1) | x | x | x | x (1) |
| A196 | | | x | | | | x |
| E27 | x(2) | x(2) | x (1) (2) | x(2) | | x(2) | x (1) (2) |
| E94 | | | | x | | | |
| PAC93 | x | x | x (1) | x | | x | x (1) |
| J93 | | x | x (1) | x | | x | x (1) |

(1) without IP67

(2) with adapter